


12

CLINICAL CONNECTIONS & SETTINGS GUIDE

I2 controls operation of Motion Control powered wrist flexion/extension with ProHand or ProETD/ETD2 open/close.

PART 1: PHYSICAL CONNECTIONS



PART 2: SOFTWARE SETTINGS

WRISTS

MOTION CONTROL POWERED FLEXION WRIST

Wrist Direction: A

RELAX DELAY

- Relax Timer: 0.1

SWITCH CONTROL

- Switch Rates A: 75
- Switch Rates B: 75
- Switch Window: 0.1
- Fast Access: On
- Co-Contraction: Off
- Auto-Switch Back: No

POWERED FLEXION WRIST SETTINGS

- Motor Speed: High
- Motor Brake: On
- Home Position Delay: 0.5

HANDS

MOTION CONTROL PROHAND, PROETD/ETD2

HAND SET UP

- Input Type: EMG
- Hand Filtering: Quick
- Channel: Dual
- Control: Differential
- Hand Direction: A

Thresholds/Outputs

- A: 15
- B: 15

Input Gains

- A: 5
- B: 5